

Mycosep® 224 AflaZON

Clean-up column
Item no. 10001951

Intended use:

Mycosep® 224 AflaZON clean-up columns are intended for the analysis of aflatoxins B1, B2, G1, G2 and zearalenone in various commodities.

Commodities studied:

Grains, feed

Recovery:

> 90 %

Variability:

< 5 %

Time for clean-up:

30 seconds

Solvents:

Acetonitrile, water, acetic acid (all HPLC grade)

Romer Labs recommends the use of Biopure™ isotope labeled internal standards.


Storage: Always store at room temperature (approx. 15 – 25 °C, 59 – 77 °F) in a dry place (preferably sealed in the original packaging protected from direct light). Do not use the clean-up columns beyond the expiration date.

IMPORTANT: Download certificate of analysis by scanning the QR code on the external label or by visiting <https://www.romerlabs.com/en/customer-resources/>

Mycotoxins



- 1 Weigh 25 g of a representative ground sample
- 2 Add 100 mL 84/16 (v/v) acetonitrile/water
- 3 Blend for 3 min or shake for 30 min
- 4 Filter or centrifuge supernatant
- 5 Transfer 5 mL to the glass tube, acidify with 50 µL acetic acid (for zearalenone)
- 6 Push through MycoSep® 224
- 7 Remove 2 mL, evaporate to near dryness
- 8 Re-dissolve in 400 µL mobile phase
- 9 Inject



MycoSep[®] 224 AflaZON

Clean-up column
Item no. 10001951

Mycotoxins



Special notes for column use:

- The packing in the columns may shift or settle during the shipping process. If this occurs, before using the column, push the frit firmly onto the packing material using a blunt object such as an unsharpened wooden pencil.
- After transferring the sample into the culture tube, place the column firmly into the top portion of the tube. You may tilt the tube slightly to wet the sides of the culture tube with the sample extract. This makes it easier to push the column into the tube.
- MycoSep[®] columns are designed for single use only. Any improper use or reuse of the columns may affect the reliability of the results.
- Culture tubes used for most MycoSep[®] columns are 15 x 85 mm (12 x 75 mm tubes are used for 112, 113 and 150 MycoSep[®] columns).
- Application notes for the use of MycoSep[®] columns that have been developed by Romer Labs are available on request. It is the responsibility of the user to validate any other methodology for the use of MycoSep[®] columns.

Disclaimer/Warranty:

The user or buyer assumes all risk in using Romer Labs products and services. Romer Labs makes no warranty of any kind, either expressed or implied, except that the materials from which its products are made are of standard quality. If any materials are defective, Romer Labs will, at its own discretion and option, repair or replace any product or components, or repeat services, which are proven to be defective in workmanship or material within its warranty periods. There is no warranty of merchantability of the products, or of the fitness of the product for any purpose. Romer Labs shall not be liable for any damages, including special or consequential damage, lost profits or expenses arising directly or indirectly from the use of its products or services. Excepting the aforementioned, Romer Labs hereby disclaims all other remedies, warranties, guarantees or liabilities, expressed or implied, arising by law or otherwise.

You can find worldwide contact information and learn more about our complete line of products for mycotoxin testing on our website.